

3701:1-58-56 **Surveys of patients and human research subjects treated with a remote afterloader unit.**

- (A) Before releasing a patient or a human research subject from licensee control, a licensee shall survey the patient or the human research subject and the remote afterloader unit with a portable radiation detection survey instrument to confirm that the source(s) has been removed from the patient or human research subject and returned to the safe shielded position.
- (B) A licensee shall retain a record of these surveys in accordance with rule 3701:1-58-87 of the Administrative Code.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

\_\_\_\_\_  
Certification

06/12/2015  
\_\_\_\_\_  
Date

Promulgated Under: 119.03  
Statutory Authority: 3748.02, 3748.04  
Rule Amplifies: 3748.04  
Prior Effective Dates: 8/15/2005