3701:1-58-69 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

(A) A licensee shall have each teletherapy unit and gamma stereotactic radiosurgery unit fully inspected and serviced during source replacement or at intervals not to exceed five years, whichever comes first, to assure proper functioning of the source exposure mechanism.

(B) This inspection and servicing may only be performed by persons specifically licensed to do so by the director, the United States nuclear regulatory commission, or an agreement state.

(C) A licensee shall keep a record of the inspection and servicing in accordance with rule 3701:1-58-100 of the Administrative Code.

Effective: 08/10/2015

Five Year Review (FYR) Dates: 05/26/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

_____
Certification

07/31/2015

_____
Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005, 10/4/10