

3701:1-58-96 **Records of periodic spot-checks for remote afterloader units.**

- (A) A licensee shall retain a record of each spot-check for remote afterloader units required by rule 3701:1-58-65 of the Administrative Code for three years.
- (B) The record must include, as applicable:
 - (1) The date of the spot-check;
 - (2) The manufacturer's name, model number, and serial number for the remote afterloader unit and source;
 - (3) An assessment of timer accuracy;
 - (4) Notations indicating the operability of each entrance door electrical interlock, radiation monitors, source exposure indicator lights, viewing and intercom systems, and clock and decayed source activity in the unit's computer; and
 - (5) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.
- (C) A licensee shall retain a copy of the procedures required by paragraph (B) of rule 3701:1-58-65 of the Administrative Code until the licensee no longer possesses the remote afterloader unit.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3743.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005